

Specification of Wrist Strap Tester/ Station



Meter alone – ST250030

Complete kit –ST250050

A unique checker that measures the resistance of wrist straps and if wall mounted heel straps, at between 750 kilohms and 10 meg ohms, meeting EN IEC 61340 specifications.

When testing, the operator simply connects to the test via the 4mm banana plug socket. To test make finger contact with the stainless push button, and “press to test”. A green LED will illuminate for pass or the red LED will illuminate for fail. An audible buzzer will sound if the red fail LED illuminates. If the resistance through the coil cord wrist band and person exceeds 35 meg ohms the red fail LED will light, if the resistance is below 750 kilohms the red fail LED will light. If the 9 volt PP3 battery falls below 6.5 volts during test the yellow battery low LED will light. The product is CE approved.

Specifications:

Power Supply	9 volt PP3 alkaline battery.
Temperature Range	Operating 5°C to 49°C (40°F to 120°F) Storage -15°C to +60°C
Relative Humidity	0% to 90% (non-condensing)
Accuracy	+/- 10%
Repeatability	+/- 10%
Weight	350g
Dimensions of Tester	130mm x 70mm x 25mm

The information contained within this spec sheet is for guidance only. We make no warranties expressed or implied and assume no liability regarding any use of this information.